



Republic of the Philippines

CIVIL AVIATION AUTHORITY OF THE PHILIPPINES

BID BULLETIN NO. 01
July 28 2025

I. Attention is hereby invited to the Bidders of the project **PURCHASE/INSTALLATION OF METEOROLOGICAL INSTRUMENTS FOR ANTIQUE AIRPORT (DUAL SITE) – Bid 25-53-07 BRAVO**.

II. Please note the following amendments to the project requirements.

Reference	Clause/Item	Change Type	Amendment		
V. SCC for GCC	Clause 2.2	Addition	This is a turn-key project; payment will be made after the issuance of the project completion certificate by CAAP.		
VI. Schedule of Requirements	Subsection V: Power Cable Provision and Civil Works	Revision	Power Cable Provision and Civil Works	Qty	Unit
			Direct Earth Burial (DEB) power cable; Civil works for the installation of DEB power cable: Excavation Jobs; Cable-laying; Sand bedding; Backfilling/compacting Jobs; 4-inch Yellow "CAUTION" Tape;	1	lot
VII. Technical Specifications – General Requirements	A.2	Revision	Construction design drawings and installation plans shall be submitted after the receipt of Notice-to- Proceed (NTP) for approval of CAAP (design review) prior to its installation/ implementation. As-built drawings shall be submitted prior to commissioning.		
VII. Technical Specifications – Equipment Requirements	B.4.1.15	Revision	For antenna mast installations, the contractor shall supply and install a radial grounding system consisting of three (3) grounding radials spaced at 120 degrees apart around the base of the mast. The final ground resistance of the entire mast grounding system shall not exceed 5 ohms, verified through ground resistance testing after installation. Note: Where the installation of radials at 120-degree separation is deemed		

			<p>impracticable due to site-specific conditions — such as uneven terrain, rocky or compacted soil, limited space, or waterlogged areas — the Contractor shall propose a modified grounding configuration. This shall be supported by technical justification and shall achieve equivalent or improved grounding performance. Any such modifications are subject to prior approval by the Implementation Engineer or Procuring Entity (PE).</p>
--	--	--	---

- III. This bid bulletin shall be an integral part of the Bidding Documents and the same shall be enclosed in the technical bid envelop/component and shall be marked/tapped accordingly.
- IV. For the information and guidance of all concerned.


GARY M. JADIE
 End-User

Noted by:


ATTY. DANJUN G. LUCAS
 Chairperson, Bids and Awards Committee – Bravo